

JAN 11 2011
2011
3/2/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130172</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Sgt. Freddie Wells</u>	ID# <u>4246</u> ✓ Date <u>1-4-11</u> ✓
Agency <u>Alaska State Troopers</u>	Phone # <u>907-863-5111</u>
Instrument Location <u>Tok, Alaska</u>	
Alco S/N <u>X173017</u> ✓ Target Value <u>.083</u> ✓ High Pressure <u>650</u>	
Alco Test Values	
1st Alco <u>.081</u> ✓	2nd Alco <u>.082</u> ✓
Signature <u>Freddie L. Wells</u>	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 10th day of MAR, 2011

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130172

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130172

JANUARY 04, 2011

OPERATOR'S NAME:

WELLS/FREDDIE/L

OPERATOR'S NUMBER: 4246

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

D.L. #: 0000912

DEPT/AGENCY: T0KE

CASE/REPORT: VOC

TEST TYPE: U

ALCO TARGET VALUE: .003

ALCO S/N: X173017

— BREATH ANALYSIS —

.003 ADJUSTED FOR 27.74 in		
ALCO TARGET	.076	16:12
BLANK TEST	.000	16:13
INTERNAL STANDARD	VERIFIED	16:13
ALCO TU 27.74 in	.001	16:13
BLANK TEST	.000	16:14
SUBJECT SAMPLE	.000	16:14
BLANK TEST	.000	16:15
ALCO TU 27.74 in	.002	16:16
BLANK TEST	.000	16:16

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130172

JANUARY 04, 2011

TIME 16:06

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	27.74 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~